
Midmark M11 - v1.0.5
Total Cycles: 4433

Name: RITTER2
Sterilizer ID: V2313228
03/19/2025 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 74.2 F 0.0
4:00 117.8 F 0.1
6:00 125.8 F 1.5
8:00 145.0 F 4.9
10:00 170.9 F 8.3
12:00 190.6 F 12.0
14:00 207.1 F 3.5
16:00 216.1 F 6.7
18:00 227.2 F 10.4
20:00 230.8 F 8.6
22:00 238.9 F 11.3
24:00 236.5 F 9.3
26:00 249.5 F 15.5
28:00 260.5 F 21.8
30:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.4 F 29.3
2:00 272.7 F 29.8
3:00 271.8 F 29.1
4:00 272.6 F 29.4

Min 271.7 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 189.5 F 0.0
10:00 190.6 F 0.0
15:00 195.7 F 0.0
20:00 203.0 F 0.0
25:00 210.4 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:44
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

03/19/2025 09:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____