
Midmark M11 - v1.0.5
Total Cycles: 364

Name: RITTER2
Sterilizer ID: V2313228
09/15/2021 06:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 73.8 F 0.0
4:00 76.3 F 0.1
6:00 92.2 F 0.2
8:00 138.3 F 1.9
10:00 170.9 F 5.3
12:00 189.0 F 9.5
14:00 206.3 F 9.9
16:00 218.1 F 8.0
18:00 230.9 F 12.5
20:00 234.4 F 9.7
22:00 236.7 F 9.5
24:00 246.5 F 14.4
26:00 258.8 F 20.9
28:00 269.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 271.6 F 29.3
2:00 271.9 F 29.4
3:00 272.3 F 29.6
4:00 272.1 F 30.0

Min 271.3 F 29.3
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 237.9 F 0.0
10:00 229.5 F 0.0
15:00 213.5 F 0.0
20:00 214.4 F 0.0
25:00 194.2 F 0.0
30:00 183.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:60
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:06:52
CYCLE COMPLETE

09/15/2021 07:31 AM
ACCEPTED: X
REJECTED:
Operator: U01