

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2511

Name: RITTER 1  
Sterilizer ID: V2178198  
11/03/2022 12:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.2 F 0.0  
2:00 153.3 F 2.5  
4:00 178.2 F 6.1  
6:00 197.3 F 10.6  
8:00 213.6 F 3.7  
10:00 223.8 F 8.2  
12:00 238.2 F 13.0  
14:00 236.7 F 9.8  
16:00 239.4 F 8.9  
18:00 247.1 F 13.9  
20:00 259.2 F 20.5  
22:00 270.3 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.6 F 29.3  
2:00 272.0 F 28.6  
3:00 272.0 F 29.0  
4:00 272.5 F 29.2

Min 271.5 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 218.4 F 0.0  
10:00 237.9 F 0.0  
15:00 236.8 F 0.0  
20:00 226.2 F 0.0  
25:00 221.7 F 0.0  
30:00 221.2 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 22:55  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:08  
CYCLE COMPLETE

11/03/2022 01:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_