

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2437

Name: RITTER2  
Sterilizer ID: V2313228  
05/30/2023 02:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
2:00 205.0 F 3.9  
4:00 215.1 F 9.5  
6:00 222.4 F 7.7  
8:00 234.0 F 11.5  
10:00 235.5 F 9.5  
12:00 237.1 F 9.1  
14:00 249.3 F 15.3  
16:00 261.5 F 22.1  
18:00 272.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.4 F 29.5  
2:00 271.6 F 28.8  
3:00 272.4 F 29.7  
4:00 272.2 F 29.2  
  
Min 271.3 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 233.7 F 0.0  
10:00 231.5 F 0.0  
15:00 230.0 F 0.0  
20:00 230.5 F 0.0  
25:00 227.0 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 2:41  
HEATING: 18:31  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

05/30/2023 02:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_