
Midmark M11 - v1.0.5
Total Cycles: 1880

Name: RITTER2
Sterilizer ID: V2313228
12/22/2022 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 197.5 F 3.8
4:00 201.5 F 9.5
6:00 222.1 F 6.9
8:00 233.7 F 11.4
10:00 235.4 F 9.6
12:00 237.7 F 9.0
14:00 249.6 F 15.8
16:00 262.7 F 23.3
18:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.8 F 29.8
2:00 271.6 F 29.1
3:00 272.8 F 30.0
4:00 272.2 F 29.4

Min 271.6 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 208.2 F 0.0
10:00 227.2 F 0.0
15:00 229.8 F 0.0
20:00 228.1 F 0.0
25:00 227.4 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 3:03
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

12/22/2022 12:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____