

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3499

Name: RITTER2  
Sterilizer ID: V2313228  
04/05/2024 09:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.9 F 0.0  
2:00 150.5 F 1.6  
4:00 168.6 F 4.6  
6:00 191.7 F 8.2  
8:00 207.7 F 12.9  
10:00 223.1 F 6.7  
12:00 228.7 F 11.3  
14:00 232.1 F 8.1  
16:00 242.5 F 12.7  
18:00 244.0 F 12.6  
20:00 255.6 F 18.8  
22:00 266.3 F 25.4  
24:00 272.8 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.5 F 29.7  
2:00 272.1 F 30.0  
3:00 272.0 F 29.4  
4:00 272.6 F 30.0

Min 271.4 F 29.2  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 232.7 F 0.0  
10:00 226.4 F 0.0  
15:00 225.4 F 0.0  
20:00 204.8 F 0.0  
25:00 189.5 F 0.0  
30:00 175.2 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:03:20  
CYCLE COMPLETE

04/05/2024 11:02 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_