

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 000301  
14:36 13/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:55  
PRESSURIZING 3:56  
STERILIZING 5:11  
136.4 C 232kPa 5:11  
136.2 C 232kPa 5:41  
136.3 C 232kPa 6:11  
136.4 C 233kPa 6:41  
136.1 C 229kPa 7:11  
136.3 C 232kPa 7:41  
136.2 C 232kPa 8:11  
136.4 C 232kPa 8:41  
136.4 C 232kPa 8:41  
VENTING 8:41  
AIR DRYING 9:41  
CYCLE COMPLETE 9:56  
-----

Digital Signature #  
266DA2F09B229462  
-----