

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3135

Name: RITTER2  
Sterilizer ID: V2313228  
11/16/2023 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.3 F 0.0  
2:00 145.4 F 2.1  
4:00 168.0 F 5.9  
6:00 192.1 F 10.7  
8:00 217.4 F 4.8  
10:00 222.4 F 10.0  
12:00 233.0 F 8.4  
14:00 241.1 F 13.1  
16:00 242.8 F 12.2  
18:00 255.9 F 19.1  
20:00 267.7 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.7 F 29.7  
2:00 271.4 F 29.0  
3:00 271.9 F 29.8  
4:00 272.3 F 29.5

Min 271.1 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
5:00 208.0 F 0.0  
10:00 200.9 F 0.0  
15:00 198.0 F 0.0  
20:00 196.7 F 0.0  
25:00 196.3 F 0.0  
30:00 197.4 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 21:20  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

11/16/2023 03:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_