

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2020

Name: RITTER 1  
Sterilizer ID: V2178198  
06/13/2022 07:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.0 F 0.0  
2:00 76.9 F 0.0  
4:00 93.6 F 0.2  
6:00 124.9 F 2.7  
8:00 151.6 F 5.9  
10:00 175.2 F 9.8  
12:00 188.4 F 8.0  
14:00 212.1 F 4.8  
16:00 223.6 F 8.6  
18:00 234.5 F 12.7  
20:00 233.6 F 8.6  
22:00 242.6 F 11.8  
24:00 243.5 F 12.0  
26:00 256.2 F 18.6  
28:00 267.2 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 271.6 F 28.2  
2:00 271.6 F 28.2  
3:00 271.7 F 28.1  
4:00 271.8 F 28.1

Min 271.6 F 28.1  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 236.4 F 0.0  
5:00 222.0 F 0.0  
Dry Cycle Aborted

FILLING: 0:58  
HEATING: 29:20  
STERILIZING: 4:00  
VENTING: 2:43  
DRYING: 8:19  
TOTAL CYCLE: 00:45:21

06/13/2022 08:25 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_