
Midmark M11 - v1.0.0
Total Cycles: 3112

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2023 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.2 F 0.0
2:00 189.6 F 0.0
4:00 198.8 F 3.3
6:00 205.6 F 8.2
8:00 221.9 F 13.1
10:00 228.4 F 8.4
12:00 240.4 F 13.4
14:00 239.9 F 11.1
16:00 237.7 F 9.5
18:00 251.7 F 16.2
20:00 263.0 F 22.7
22:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.0
2:00 271.4 F 28.3
3:00 272.5 F 29.2
4:00 272.0 F 28.7

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 208.5 F 0.0
10:00 214.6 F 0.0
15:00 211.6 F 0.0
20:00 215.2 F 0.0
25:00 216.9 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

04/14/2023 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____