
Midmark M11 - v1.0.0
Total Cycles: 2877

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2023 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 148.4 F 2.9
4:00 172.9 F 6.8
6:00 197.4 F 11.6
8:00 211.0 F 5.3
10:00 230.9 F 10.2
12:00 228.5 F 7.5
14:00 241.7 F 12.2
16:00 239.6 F 10.3
18:00 252.9 F 16.8
20:00 264.5 F 23.6
22:00 270.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.8 F 29.1
2:00 271.5 F 28.7
3:00 272.5 F 29.1
4:00 271.5 F 28.6

Min 271.5 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 238.7 F 0.0
10:00 222.8 F 0.0
15:00 232.0 F 0.0
20:00 226.8 F 0.0
25:00 218.9 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:05
CYCLE COMPLETE

02/08/2023 03:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____