
Midmark M11 - v1.0.0
Total Cycles: 1837

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2022 06:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.9 F 0.0
2:00 81.6 F 0.0
4:00 97.7 F 0.0
6:00 125.9 F 0.0
8:00 161.4 F 2.5
10:00 179.2 F 5.9
12:00 204.5 F 9.2
14:00 220.7 F 12.6
16:00 220.6 F 5.3
18:00 232.4 F 9.1
20:00 242.1 F 13.1
22:00 236.5 F 9.5
24:00 239.3 F 9.6
26:00 245.9 F 13.1
28:00 257.2 F 19.2
30:00 267.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.6 F 29.0
2:00 271.2 F 28.2
3:00 272.5 F 29.2
4:00 272.2 F 28.6

Min 271.0 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 228.3 F 0.0
10:00 196.7 F 0.0
15:00 209.6 F 0.0
20:00 211.2 F 0.0
25:00 223.7 F 0.0
30:00 241.9 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:21
CYCLE COMPLETE

04/19/2022 08:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____