

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002374  
11:47 06/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:14  
STERILIZING 4:21  
136.5 C 223kPa 4:21  
136.3 C 224kPa 4:51  
136.1 C 220kPa 5:21  
136.3 C 222kPa 5:51  
136.4 C 224kPa 6:21  
136.1 C 221kPa 6:51  
136.2 C 222kPa 7:21  
136.4 C 222kPa 7:51  
136.4 C 222kPa 7:51  
VENTING 7:51  
AIR DRYING 8:45  
CYCLE COMPLETE 8:48

-----  
Digital Signature #  
462E3DA91452D2A0  
-----