
Midmark M11 - v1.0.5
Total Cycles: 3626

Name: RITTER2
Sterilizer ID: V2313228
05/16/2024 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.1 F 0.0
2:00 118.5 F 1.0
4:00 132.9 F 3.4
6:00 160.9 F 6.6
8:00 181.6 F 10.1
10:00 206.3 F 8.7
12:00 215.0 F 7.2
14:00 225.0 F 11.3
16:00 229.3 F 7.3
18:00 239.8 F 12.4
20:00 238.7 F 10.6
22:00 250.5 F 16.4
24:00 260.8 F 22.4
26:00 270.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 272.4 F 30.3
2:00 271.8 F 29.7
3:00 271.1 F 29.7
4:00 272.4 F 30.3

Min 271.1 F 29.4
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 203.9 F 0.0
10:00 202.7 F 0.0
15:00 217.4 F 0.0
20:00 223.6 F 0.0
25:00 222.6 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:53
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:05:27
CYCLE COMPLETE

05/16/2024 12:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____