

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4551

Name: RITTER 1  
Sterilizer ID: V2178198  
05/14/2024 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.0 F 0.0  
2:00 143.6 F 2.0  
4:00 153.6 F 4.9  
6:00 175.9 F 8.4  
8:00 196.5 F 12.5  
10:00 212.8 F 5.7  
12:00 224.3 F 9.9  
14:00 231.7 F 9.0  
16:00 238.6 F 10.7  
18:00 234.9 F 8.2  
20:00 248.5 F 14.7  
22:00 259.6 F 20.8  
24:00 270.0 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.9 F 29.5  
2:00 271.5 F 29.0  
3:00 272.0 F 29.0  
4:00 272.9 F 29.4

Min 270.9 F 28.4  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 213.6 F 0.0  
10:00 235.5 F 0.0  
15:00 230.7 F 0.0  
20:00 209.2 F 0.0  
25:00 213.5 F 0.0  
30:00 213.2 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 24:58  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:41  
CYCLE COMPLETE

05/14/2024 12:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_