
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001914
7:41 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:45
STERILIZING 3:52
136.3 C 221kPa 3:52
136.3 C 223kPa 4:22
136.2 C 224kPa 4:52
136.4 C 222kPa 5:22
136.4 C 223kPa 5:52
135.9 C 222kPa 6:22
136.0 C 221kPa 6:52
136.4 C 225kPa 7:22
136.4 C 225kPa 7:22
VENTING 7:22
AIR DRYING 8:09
CYCLE COMPLETE 28:00

Digital Signature #
E2BA9E056BE91733
