
Midmark M11 - v1.0.0
Total Cycles: 5042

Name: RITTER 1
Sterilizer ID: V2178198
10/31/2024 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.2 F 0.0
2:00 92.3 F 0.0
4:00 109.6 F 0.2
6:00 128.2 F 1.0
8:00 161.7 F 2.9
10:00 182.8 F 5.8
12:00 198.7 F 9.0
14:00 216.3 F 12.8
16:00 221.2 F 6.4
18:00 231.4 F 10.7
20:00 233.7 F 7.7
22:00 242.0 F 12.2
24:00 239.6 F 10.3
26:00 251.9 F 16.2
28:00 262.5 F 22.3
30:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.0 F 29.1
2:00 272.7 F 29.4
3:00 272.0 F 28.8
4:00 271.7 F 29.1

Min 271.5 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 165.8 F 0.0
10:00 198.9 F 0.0
15:00 198.4 F 0.0
20:00 202.5 F 0.0
25:00 216.1 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:16
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

10/31/2024 02:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____