
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002063
8:40 19/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 4:14
STERILIZING 5:12
136.4 C 220kPa 5:12
136.4 C 225kPa 5:42
136.4 C 223kPa 6:12
136.2 C 223kPa 6:42
136.4 C 222kPa 7:12
136.4 C 223kPa 7:42
136.2 C 222kPa 8:12
136.4 C 222kPa 8:42
136.4 C 222kPa 8:42
VENTING 8:42
AIR DRYING 9:37
CYCLE COMPLETE 9:42

Digital Signature #
AC18E10E66661312
