
Midmark M11 - v1.0.0
Total Cycles: 4834

Name: RITTER 1
Sterilizer ID: V2178198
08/20/2024 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 73.7 F 0.0
4:00 87.1 F 0.1
6:00 123.7 F 0.3
8:00 151.6 F 0.9
10:00 179.5 F 2.0
12:00 195.9 F 5.0
14:00 209.5 F 8.2
16:00 220.3 F 11.7
18:00 221.9 F 5.3
20:00 232.2 F 9.7
22:00 238.1 F 10.0
24:00 241.2 F 11.3
26:00 237.3 F 9.1
28:00 250.5 F 15.4
30:00 261.3 F 21.5
32:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.7 F 28.9
2:00 271.7 F 28.0
3:00 272.6 F 28.9
4:00 272.1 F 28.4

Min 271.6 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 202.8 F 0.0
10:00 209.6 F 0.0
15:00 220.1 F 0.0
20:00 218.3 F 0.0
25:00 219.6 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 32:40
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:52
CYCLE COMPLETE

08/20/2024 09:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____