
Midmark M11 - v1.0.0
Total Cycles: 4170

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2023 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.2 F 0.0
2:00 99.6 F 0.5
4:00 112.7 F 2.1
6:00 146.4 F 4.8
8:00 164.1 F 8.2
10:00 189.9 F 13.0
12:00 217.6 F 5.3
14:00 228.0 F 9.8
16:00 234.0 F 8.9
18:00 240.5 F 11.3
20:00 237.0 F 9.2
22:00 250.4 F 15.6
24:00 261.8 F 22.1
26:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 271.9 F 28.5
2:00 272.2 F 28.7
3:00 272.4 F 29.0
4:00 272.0 F 28.9

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 238.7 F 0.0
10:00 216.7 F 0.0
15:00 216.0 F 0.0
20:00 204.4 F 0.0
25:00 213.2 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 26:14
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

12/29/2023 01:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____