

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5425

Name: RITTER 1  
Sterilizer ID: V2178198  
03/20/2025 08:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 78.5 F 0.4  
4:00 87.7 F 1.8  
6:00 124.2 F 4.2  
8:00 155.1 F 7.2  
10:00 180.2 F 10.5  
12:00 197.6 F 4.9  
14:00 210.5 F 4.6  
16:00 221.3 F 8.1  
18:00 231.3 F 12.0  
20:00 229.8 F 7.4  
22:00 242.2 F 12.5  
24:00 239.0 F 10.2  
26:00 251.7 F 16.4  
28:00 262.6 F 22.8  
30:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.4 F 29.1  
2:00 271.8 F 28.8  
3:00 272.4 F 28.9  
4:00 271.9 F 28.8  
  
Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
5:00 200.3 F 0.0  
10:00 209.0 F 0.0  
15:00 204.1 F 0.0  
20:00 206.3 F 0.0  
25:00 208.2 F 0.0  
30:00 208.6 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 30:18  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:08:35  
CYCLE COMPLETE

03/20/2025 09:28 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_