
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001276
12:13 28/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:02
PRESSURIZING 4:03
STERILIZING 5:21
136.0 C 220kPa 5:21
136.4 C 224kPa 5:51
136.2 C 224kPa 6:21
136.4 C 225kPa 6:51
136.4 C 224kPa 7:21
136.2 C 224kPa 7:51
136.4 C 224kPa 8:21
136.4 C 224kPa 8:51
136.4 C 224kPa 8:51
VENTING 8:51
AIR DRYING 9:47
CYCLE COMPLETE 19:26

Digital Signature #
7D6C72F2DC248464
