

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2561

Name: RITTER 1  
Sterilizer ID: V2178198  
11/15/2022 12:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.5 F 0.0  
2:00 171.8 F 3.0  
4:00 177.0 F 7.1  
6:00 214.9 F 12.5  
8:00 224.1 F 6.7  
10:00 235.4 F 11.9  
12:00 234.7 F 8.7  
14:00 242.1 F 9.9  
16:00 246.6 F 13.7  
18:00 259.2 F 20.6  
20:00 270.6 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.4 F 29.3  
2:00 272.2 F 28.7  
3:00 272.2 F 29.1  
4:00 272.6 F 29.0  
  
Min 271.6 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 209.0 F 0.0  
10:00 220.6 F 0.0  
15:00 223.7 F 0.0  
20:00 225.0 F 0.0  
25:00 229.5 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 20:39  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:09  
CYCLE COMPLETE

11/15/2022 01:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_