
Midmark M11 - v1.0.5
Total Cycles: 1252

Name: RITTER2
Sterilizer ID: V2313228
07/04/2022 04:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.3 F 0.0
2:00 161.9 F 2.8
4:00 178.3 F 7.0
6:00 202.1 F 12.2
8:00 216.6 F 7.1
10:00 231.9 F 12.3
12:00 232.8 F 9.5
14:00 237.4 F 9.2
16:00 245.9 F 14.3
18:00 258.0 F 20.9
20:00 269.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 272.0 F 29.5
2:00 272.2 F 29.7
3:00 272.2 F 29.9
4:00 272.0 F 30.2

Min 271.5 F 29.5
Max 272.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 204.4 F 0.0
10:00 215.4 F 0.0
15:00 218.5 F 0.0
20:00 217.2 F 0.0
25:00 216.3 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

07/04/2022 05:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____