
Midmark M11 - v1.0.0
Total Cycles: 4257

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2024 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 175.7 F 1.7
4:00 193.9 F 5.5
6:00 206.0 F 10.9
8:00 222.2 F 5.0
10:00 230.7 F 10.2
12:00 234.1 F 7.9
14:00 244.8 F 12.8
16:00 242.2 F 11.3
18:00 255.1 F 17.9
20:00 266.4 F 24.7
22:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.6
2:00 272.8 F 29.1
3:00 271.7 F 28.7
4:00 272.4 F 28.7

Min 271.5 F 28.4
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 212.6 F 0.0
10:00 203.5 F 0.0
15:00 203.5 F 0.0
20:00 202.3 F 0.0
25:00 240.9 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

02/01/2024 11:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____