
Midmark M11 - v1.0.5
Total Cycles: 96

Name: RITTER2
Sterilizer ID: V2313228
06/14/2021 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.8 F 0.0
2:00 135.5 F 1.7
4:00 154.0 F 4.9
6:00 182.5 F 9.3
8:00 203.2 F 7.3
10:00 217.0 F 8.5
12:00 230.7 F 11.9
14:00 238.3 F 11.6
16:00 238.6 F 10.5
18:00 252.0 F 17.0
20:00 263.3 F 23.7
22:00 271.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.9
1:00 272.9 F 30.5
2:00 271.9 F 30.0
3:00 272.7 F 30.6
4:00 272.2 F 29.9

Min 271.7 F 29.8
Max 273.2 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 222.4 F 0.0
10:00 227.5 F 0.0
15:00 222.4 F 0.0
20:00 222.2 F 0.0
25:00 220.3 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

06/14/2021 02:21 PM
ACCEPTED: X
REJECTED:
Operator: U01