
Midmark M11 - v1.0.0
Total Cycles: 4673

Name: RITTER 1
Sterilizer ID: V2178198
06/25/2024 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 72.4 F 0.0
4:00 95.1 F 0.4
6:00 119.6 F 1.6
8:00 148.3 F 4.1
10:00 172.3 F 7.5
12:00 189.7 F 10.8
14:00 203.1 F 4.0
16:00 214.6 F 5.8
18:00 224.7 F 9.2
20:00 234.7 F 13.2
22:00 235.1 F 9.2
24:00 240.2 F 10.5
26:00 244.6 F 12.5
28:00 256.4 F 18.7
30:00 267.0 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.6 F 28.1
2:00 271.8 F 28.3
3:00 272.2 F 28.5
4:00 272.5 F 28.8

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 201.7 F 0.0
10:00 203.9 F 0.0
15:00 217.4 F 0.0
20:00 219.1 F 0.0
25:00 223.7 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:09:31
CYCLE COMPLETE

06/25/2024 10:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____