
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000882
14:18 26/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:31
STERILIZING 4:43
136.0 C 227kPa 4:43
136.4 C 228kPa 5:13
136.4 C 231kPa 5:43
136.2 C 230kPa 6:13
136.4 C 230kPa 6:43
136.3 C 231kPa 7:13
136.1 C 228kPa 7:43
136.2 C 230kPa 8:13
136.2 C 230kPa 8:13
VENTING 8:13
AIR DRYING 9:08
CYCLE COMPLETE 20:14

Digital Signature #
5744F852BA2AF6E6
