
Midmark M11 - v1.0.0
Total Cycles: 586

Name: MIDMARK
Sterilizer ID: V2178198
12/29/2020 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.8 F 0.0
2:00 118.8 F 0.1
4:00 143.3 F 1.9
6:00 178.0 F 4.7
8:00 198.1 F 9.3
10:00 221.2 F 8.0
12:00 221.9 F 7.3
14:00 234.5 F 12.1
16:00 235.9 F 8.9
18:00 241.3 F 10.5
20:00 247.0 F 13.3
22:00 259.0 F 19.8
24:00 269.9 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.1 F 28.4
2:00 271.3 F 28.1
3:00 272.2 F 28.4
4:00 271.2 F 27.8

Min 271.2 F 27.7
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 215.7 F 0.0
10:00 207.6 F 0.0
15:00 213.5 F 0.0
20:00 221.4 F 0.0
25:00 238.2 F 0.0
30:00 235.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

12/29/2020 02:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____