
Midmark M11 - v1.0.0
Total Cycles: 4095

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2023 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 83.4 F 0.0
2:00 84.1 F 0.4
4:00 87.5 F 1.6
6:00 109.3 F 3.6
8:00 142.0 F 6.1
10:00 179.5 F 9.1
12:00 195.1 F 12.6
14:00 205.5 F 3.3
16:00 211.4 F 6.3
18:00 221.5 F 9.7
20:00 232.9 F 12.4
22:00 234.8 F 9.4
24:00 239.9 F 10.6
26:00 244.0 F 12.5
28:00 254.2 F 17.7
30:00 263.7 F 23.3
32:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.2
2:00 271.9 F 28.7
3:00 271.2 F 28.5
4:00 272.3 F 29.2

Min 271.1 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 217.2 F 0.0
10:00 212.3 F 0.0
15:00 211.8 F 0.0
20:00 209.4 F 0.0
25:00 224.8 F 0.0
30:00 234.3 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 32:26
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:11:11
CYCLE COMPLETE

12/07/2023 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____