

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1503

Name: RITTER2  
Sterilizer ID: V2313228  
09/19/2022 01:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
2:00 180.7 F 3.2  
4:00 190.0 F 7.3  
6:00 209.8 F 12.4  
8:00 220.0 F 7.6  
10:00 234.7 F 13.0  
12:00 237.2 F 10.7  
14:00 236.0 F 9.3  
16:00 250.1 F 16.1  
18:00 262.1 F 22.9  
20:00 272.7 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 29.7  
1:00 272.3 F 30.3  
2:00 272.8 F 30.2  
3:00 272.0 F 29.6  
4:00 272.3 F 30.4

Min 271.8 F 29.6  
Max 272.9 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 220.7 F 0.0  
10:00 233.2 F 0.0  
15:00 236.5 F 0.0  
20:00 226.0 F 0.0  
25:00 222.9 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 21:11  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:56  
CYCLE COMPLETE

09/19/2022 02:47 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_