
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006305
8:55 15/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:47
STERILIZING 4:56
136.4 C 223kPa 4:56
136.1 C 222kPa 5:26
136.3 C 222kPa 5:56
136.4 C 218kPa 6:26
136.0 C 221kPa 6:56
136.1 C 218kPa 7:26
136.2 C 218kPa 7:56
136.4 C 222kPa 8:26
136.4 C 222kPa 8:26
VENTING 8:26
AIR DRYING 9:12
CYCLE COMPLETE 39:12

Digital Signature #
7476D59514BAEB92
