
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007320
7:46 06/05/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:48
STERILIZING 5:02
136.4 C 220kPa 5:02
136.1 C 220kPa 5:32
136.2 C 221kPa 6:02
136.2 C 221kPa 6:32
136.4 C 222kPa 7:02
136.4 C 221kPa 7:32
136.4 C 220kPa 8:02
136.1 C 221kPa 8:32
136.1 C 221kPa 8:32
VENTING 8:32
AIR DRYING 9:30
CYCLE COMPLETE 9:36

Digital Signature #
D13F74C48E33EAAC
