
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004121
10:55 27/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:56
STERILIZING 3:04
136.4 C 221kPa 3:04
136.3 C 223kPa 3:35
136.4 C 224kPa 4:05
136.3 C 224kPa 4:35
136.1 C 221kPa 5:05
136.4 C 224kPa 5:35
136.4 C 224kPa 6:05
136.4 C 224kPa 6:35
136.4 C 224kPa 6:35
VENTING 6:35
AIR DRYING 7:24
CYCLE COMPLETE 12:56

Digital Signature #
0C376044D75650F4
