
Midmark M11 - v1.0.0
Total Cycles: 1871

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2022 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 179.5 F 3.2
4:00 198.1 F 7.4
6:00 222.7 F 13.0
8:00 227.3 F 7.6
10:00 239.0 F 13.0
12:00 238.5 F 10.2
14:00 235.8 F 8.1
16:00 250.2 F 15.3
18:00 262.2 F 22.1
20:00 272.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.4 F 29.1
4:00 272.1 F 28.7

Min 271.2 F 28.1
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 226.5 F 0.0
10:00 255.7 F 0.0
15:00 233.0 F 0.0
20:00 220.2 F 0.0
25:00 217.3 F 0.0
30:00 174.6 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:21
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

04/26/2022 12:33 PM
ACCEPTED: X
REJECTED:
Operator: U02