
Midmark M11 - v1.0.0
Total Cycles: 2598

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 174.9 F 0.8
4:00 193.7 F 4.6
6:00 231.2 F 9.6
8:00 230.0 F 7.6
10:00 237.9 F 10.9
12:00 235.1 F 8.4
14:00 240.4 F 10.0
16:00 248.8 F 14.3
18:00 261.7 F 21.6
20:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.6 F 28.7
2:00 271.3 F 27.9
3:00 272.5 F 28.9
4:00 272.2 F 28.3

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 245.4 F 0.0
10:00 230.2 F 0.0
15:00 233.0 F 0.0
20:00 223.0 F 0.0
25:00 225.7 F 0.0
Dry Cycle Aborted

FILLING: 1:15
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:12
DRYING: 27:57
TOTAL CYCLE: 00:56:43

11/22/2022 05:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____