
Midmark M11 - v1.0.5
Total Cycles: 3680

Name: RITTER2
Sterilizer ID: V2313228
06/10/2024 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 70.9 F 0.0
4:00 83.3 F 0.5
6:00 96.3 F 1.9
8:00 127.0 F 4.2
10:00 152.8 F 6.9
12:00 173.7 F 9.8
14:00 189.0 F 12.9
16:00 205.6 F 4.0
18:00 215.3 F 7.2
20:00 227.6 F 10.7
22:00 228.3 F 8.2
24:00 236.1 F 11.3
26:00 234.6 F 8.9
28:00 246.6 F 14.5
30:00 256.9 F 19.9
32:00 266.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 271.8 F 29.9
2:00 273.0 F 30.3
3:00 273.1 F 30.3
4:00 271.8 F 29.9

Min 271.0 F 29.2
Max 273.3 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 196.6 F 0.0
10:00 183.2 F 0.0
15:00 176.0 F 0.0
20:00 172.7 F 0.0
25:00 172.4 F 0.0
30:00 172.6 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 33:59
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:12:27
CYCLE COMPLETE

06/10/2024 10:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____