

-----  
Midmark M11 - v1.0.5  
Total Cycles: 399

Name: RITTER2  
Sterilizer ID: V2313228  
09/28/2021 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.5 F 0.0  
2:00 112.4 F 0.8  
4:00 117.3 F 2.7  
6:00 147.2 F 6.0  
8:00 179.7 F 10.3  
10:00 198.6 F 5.0  
12:00 213.8 F 8.2  
14:00 227.6 F 13.0  
16:00 231.5 F 9.8  
18:00 235.6 F 9.1  
20:00 245.0 F 14.0  
22:00 256.3 F 20.2  
24:00 267.0 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 30.2  
1:00 272.2 F 30.3  
2:00 272.4 F 30.5  
3:00 272.1 F 30.3  
4:00 272.2 F 30.3

Min 271.9 F 30.1  
Max 272.7 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.1 F 0.0  
5:00 222.3 F 0.0  
10:00 239.0 F 0.0  
15:00 231.1 F 0.0  
20:00 210.4 F 0.0  
25:00 199.9 F 0.0  
30:00 191.1 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:19  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:03:16  
CYCLE COMPLETE

09/28/2021 09:52 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01