
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005359
14:51 11/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:56
STERILIZING 2:49
136.4 C 222kPa 2:49
136.4 C 225kPa 3:19
136.1 C 223kPa 3:49
136.1 C 223kPa 4:19
136.2 C 221kPa 4:49
135.9 C 220kPa 5:19
136.3 C 224kPa 5:49
136.3 C 221kPa 6:19
136.3 C 221kPa 6:19
VENTING 6:19
AIR DRYING 7:10
CYCLE COMPLETE 14:40

Digital Signature #
EC4A917935525150
