
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001417
11:53 09/09/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 5:51
STERILIZING 7:05
136.0 C 220kPa 7:05
136.4 C 220kPa 7:35
136.2 C 217kPa 8:05
136.5 C 224kPa 8:35
136.1 C 220kPa 9:05
136.3 C 222kPa 9:35
136.4 C 223kPa 10:05
136.4 C 222kPa 10:35
VENTING 10:35
AIR DRYING 11:37
CYCLE COMPLETE 13:46

Digital Signature #
E303E0CEF914A9BA
