
Midmark M11 - v1.0.0
Total Cycles: 1180

Name: RITTER 1
Sterilizer ID: V2178198
09/16/2021 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.1 F 0.0
2:00 158.2 F 2.3
4:00 178.2 F 5.8
6:00 222.9 F 10.7
8:00 221.0 F 5.4
10:00 230.5 F 11.1
12:00 233.0 F 7.5
14:00 245.4 F 13.5
16:00 243.7 F 11.8
18:00 256.4 F 18.3
20:00 267.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.6 F 28.7
2:00 271.3 F 27.9
3:00 272.4 F 28.8
4:00 272.2 F 28.3

Min 271.0 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 242.0 F 0.0
10:00 227.7 F 0.0
15:00 226.6 F 0.0
20:00 225.6 F 0.0
25:00 228.4 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

09/16/2021 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____