
Midmark M11 - v1.0.0
Total Cycles: 4575

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2024 09:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.9 F 0.0
2:00 144.6 F 2.0
4:00 154.6 F 5.1
6:00 178.0 F 8.7
8:00 199.7 F 13.1
10:00 211.6 F 4.9
12:00 223.4 F 8.9
14:00 234.6 F 13.3
16:00 235.2 F 9.4
18:00 240.0 F 10.5
20:00 244.9 F 12.8
22:00 256.1 F 18.6
24:00 266.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.8 F 28.4
3:00 271.9 F 28.9
4:00 272.4 F 29.0

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 210.7 F 0.0
10:00 204.5 F 0.0
15:00 201.0 F 0.0
20:00 206.8 F 0.0
25:00 204.3 F 0.0
30:00 204.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:45
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:04:16
CYCLE COMPLETE

05/23/2024 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____