

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000960  
10:19 07/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:54  
PRESSURIZING 3:57  
STERILIZING 5:05  
136.0 C 229kPa 5:05  
136.4 C 233kPa 5:35  
136.3 C 232kPa 6:05  
136.4 C 233kPa 6:35  
136.4 C 233kPa 7:05  
136.1 C 230kPa 7:35  
136.3 C 233kPa 8:05  
136.4 C 233kPa 8:35  
136.4 C 233kPa 8:35  
VENTING 8:35  
AIR DRYING 9:31  
CYCLE COMPLETE 24:31

-----  
Digital Signature #  
8DE5F00BE71200ED  
-----