
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002278
11:04 18/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:44
STERILIZING 3:54
136.0 C 221kPa 3:54
136.3 C 225kPa 4:24
136.3 C 225kPa 4:54
136.1 C 223kPa 5:24
136.1 C 223kPa 5:54
136.4 C 224kPa 6:24
136.3 C 224kPa 6:54
136.3 C 225kPa 7:24
136.3 C 225kPa 7:24
VENTING 7:24
AIR DRYING 8:18
CYCLE COMPLETE 21:23

Digital Signature #
761889E2ACC512BA
