

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2398

Name: RITTER2  
Sterilizer ID: V2313228  
05/23/2023 09:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
2:00 173.3 F 3.0  
4:00 181.5 F 7.2  
6:00 207.2 F 12.4  
8:00 222.4 F 7.1  
10:00 236.3 F 12.7  
12:00 237.4 F 10.4  
14:00 236.2 F 9.1  
16:00 250.8 F 16.1  
18:00 263.0 F 23.0  
20:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.6 F 29.7  
2:00 272.1 F 29.1  
3:00 271.2 F 28.9  
4:00 272.7 F 29.8

Min 271.2 F 28.7  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 209.6 F 0.0  
10:00 222.0 F 0.0  
15:00 220.0 F 0.0  
20:00 225.3 F 0.0  
25:00 223.4 F 0.0  
30:00 220.2 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

05/23/2023 10:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_