
Midmark M11 - v1.0.5
Total Cycles: 1211

Name: RITTER2
Sterilizer ID: V2313228
06/22/2022 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 180.3 F 3.1
4:00 191.1 F 7.6
6:00 214.5 F 13.0
8:00 225.9 F 8.2
10:00 239.0 F 12.7
12:00 239.3 F 11.5
14:00 238.7 F 10.3
16:00 252.0 F 17.0
18:00 263.5 F 23.7
20:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.8 F 29.5
2:00 272.6 F 29.6
3:00 271.6 F 29.0
4:00 271.6 F 29.5

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 213.1 F 0.0
10:00 220.4 F 0.0
15:00 219.4 F 0.0
20:00 216.9 F 0.0
25:00 214.8 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

06/22/2022 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____