
Midmark M11 - v1.0.0
Total Cycles: 2900

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2023 02:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.7 F 0.0
2:00 201.1 F 4.4
4:00 209.2 F 9.6
6:00 223.4 F 6.4
8:00 233.3 F 10.2
10:00 233.3 F 7.6
12:00 244.5 F 13.2
14:00 244.6 F 12.4
16:00 257.3 F 19.2
18:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 271.9 F 28.4
3:00 271.3 F 28.3
4:00 272.4 F 29.1

Min 271.2 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 234.0 F 0.0
10:00 230.3 F 0.0
15:00 231.8 F 0.0
20:00 226.7 F 0.0
25:00 229.5 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

02/14/2023 03:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____