
Midmark M11 - v1.0.0
Total Cycles: 1928

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2022 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 181.2 F 2.3
4:00 179.7 F 5.7
6:00 196.7 F 10.3
8:00 212.8 F 4.8
10:00 228.3 F 8.8
12:00 236.7 F 10.7
14:00 239.2 F 10.9
16:00 236.6 F 8.7
18:00 250.3 F 15.3
20:00 261.9 F 21.9
22:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.8
2:00 271.7 F 28.1
3:00 272.1 F 28.8
4:00 272.4 F 28.6

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 225.1 F 0.0
10:00 233.9 F 0.0
15:00 231.9 F 0.0
20:00 230.9 F 0.0
25:00 229.9 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 22:26
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

05/11/2022 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____