
Midmark M11 - v1.0.5
Total Cycles: 2418

Name: RITTER2
Sterilizer ID: V2313228
05/25/2023 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 184.0 F 4.0
4:00 202.0 F 9.2
6:00 220.7 F 7.2
8:00 230.9 F 10.9
10:00 233.5 F 8.7
12:00 238.6 F 9.9
14:00 248.1 F 14.7
16:00 260.5 F 21.7
18:00 271.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.6
2:00 271.5 F 28.8
3:00 272.6 F 29.8
4:00 272.2 F 29.3

Min 271.3 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 230.0 F 0.0
10:00 242.6 F 0.0
15:00 236.2 F 0.0
20:00 222.5 F 0.0
25:00 221.2 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

05/25/2023 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____