

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007695  
10:45 22/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:03  
STERILIZING 3:04  
136.4 C 221kPa 3:04  
136.3 C 224kPa 3:34  
136.2 C 221kPa 4:04  
136.3 C 224kPa 4:34  
136.4 C 224kPa 5:04  
136.1 C 218kPa 5:34  
136.4 C 222kPa 6:04  
136.3 C 223kPa 6:34  
136.3 C 223kPa 6:34  
VENTING 6:34  
AIR DRYING 7:24  
CYCLE COMPLETE 37:24

-----  
Digital Signature #  
5DA383F377E7AD74  
-----